



South Coast Air Quality Management District
21865 Copley Drive, Diamond Bar, CA 91765-4178
PERMIT TO OPERATE

Page 1
Permit No.
G43810
A/N 529394

This initial permit must be renewed ANNUALLY unless the equipment is moved, or changes ownership.
If the billing for the annual renewal fee (Rule 301.f) is not received by the expiration date, contact the District.

**Legal Owner
or Operator:**

STERIGENICS US, INC.
2015 SPRING RD, STE 650
OAK BROOK, IL 60523

ID 126197

Equipment Location: 4900 S GIFFORD AVE, LOS ANGELES, CA 90058-2785

Equipment Description :

Ethylene Oxide Aeration Room, Ambient Temperature, 59'-0" L x 39'-0" W x 13'-0" H, Vented to an ETO Abator
(Catalytic Oxidizer)

Conditions :

1. Operation of this equipment shall be conducted in accordance with all data and specifications submitted with the application under which this permit is issued unless otherwise noted below.
2. This equipment shall be properly maintained and kept in good operating condition at all times.
3. This equipment shall not be operated unless it is vented to the ETO control devices that are in compliance with the AQMD Rule 1405 and have been issued permit to construct or operate by the AQMD.
4. This aeration room shall be used for degassing of ethylene oxide (ETO) of the products sterilized in this facility only.
5. This equipment and all the devices and components which are connected to this equipment shall be leak tested every six months using the latest CARB test method during conditions of maximum sterilant gas use.
6. There shall be no staging or sterilized products in an uncontrolled environment. Any test or bio indicator removal shall be conducted in enclosed environment that is vented to an ETO control equipment.
7. The operator of this equipment shall comply with all requirements specified in the Ethylene Oxide Airborne Toxic Control Measure (ATCM) for sterilizers and aerators, Parts 1 and 2 under Title 17 of California Code of Regulations, Sections 93108 and 93108.5 (17 CCR, Sections 93108 & 93108.5).
8. The operator of this equipment shall comply with all requirements specified in the National Emission Standards for Hazardous Air Pollutants (NESHAP) for ethylene oxide commercial sterilization and fumigation operations under Code of Federal Regulations, Title 40, Part 63 Subpart O (40 CFR 63, Subpart O).
9. Records of leak tests, maintenance and corrective actions, and other records required by this permit shall be maintained on file for a minimum of five years and shall be made available to the AQMD personnel upon request. At minimum the most recent two years of records shall be retained on site.



FILE COPY

South Coast Air Quality Management District
Certified Copy



South Coast Air Quality Management District
21865 Copley Drive, Diamond Bar, CA 91765-4178

Page 2
Permit No.
G43810
A/N 529394

PERMIT TO OPERATE

NOTICE

In accordance with Rule 206, this Permit to Operate or copy shall be posted on or within 8 meters of the equipment.

This permit does not authorize the emission of air contaminants in excess of those allowed by Division 26 of the Health and Safety Code of the State of California or the applicable Rules and Regulations of the South Coast Air Quality Management District (SCAQMD). This permit cannot be considered as permission to violate existing laws, ordinances, regulations or statutes of other government agencies.

Executive Officer


By Dorris M. Bailey/JY02
11/23/2016



FILE COPY
South Coast Air Quality Management District
Certified Copy